

Corrigendum

The specifications of Item No.01 advertised in RFQ No. PMU/PROC/01-15/2020-21 dated 24-11-2020 to be read as:

TECHNICAL SPECIFICATIONS

SR #	NAME & DETAIL
01	<p><u>Pulse Oximeter (Hospital Grade) (Qty: 50)</u></p> <p>Non-invasive measurement of oxygen saturation and pulse rate with LED/ LCD display. Large display screen for plethysmography waveform. Proper Desktop model. Display of Oxygen Saturation, Pulse Rate. Oxygen Saturation measurement range from 0 -100%. Pulse strength perfusion indication. Pulse Rate measurement from 20-250 bpm. Visual and audible indication of alarms. High and low alarms settings. Adult finger and pediatric sensor (reusable type). 220 V/50Hz with minimum 2 hours battery backup.</p> <p>Country of Manufacturer: USA/EU/UK/Japan Quality Certificate: FDA510K/CE MDD/ Jp MHLW Mode: DDP Manufacturers/ Sole Authorized Distributors may participate only.</p> <p>Warranty: Comprehensive 2 Years</p>